

APM CPF Series Operation Manual

for Advanced Pressure Module





Contents

Overview 1
Introduction
Operation 2
Advanced Pressure Module (APM) Instructions
Specifications 3
Communications
Interface Description
Setting up the Serial Port for Remote Control
Using Commands
Remote Commands and Error Codes
Entering Commands10
APMCAL 11
Calibration Software11
Certifications 12
Support 13
Contact Us
Warranty

Overview

INTRODUCTION

The Crystal APM CPF series of pressure modules allow you to add pressure measurement capability to your JOFRA process calibrator. The APM uses the same reliable, high accuracy, digital temperature compensated technology found in the nVision, housed in a new rugged enclosure with an integrated cable to connect to your JOFRA calibrator. In addition, you can connect the APM to a hand-held pressure calibrator to increase the range of that calibrator.

Supported JOFRA Calibrators

The Crystal APM has been engineered to work seamlessly with many of the most popular JOFRA calibrators. Below is a list of compatible JOFRA calibrators.

- AMC910
- ASC300
- ASC301
- ASC321
- ASC-400
- ASC-400 BARO
- HPC500
- HPC502
- HPC600

Note: This manual includes information on the APM CPF modules only. For details on the operation of JOFRA calibrators, please refer to their User Manuals.

What's Included

Each unit includes an APM CPF pressure module, a fitting of your choice (NPT, BSP, or M20), NIST traceable calibration certificate, and AMETEK product CD. Crystal Engineering calibration facilities are A2LA accredited, (#2601.01) which is internationally recognized by ILAC.

Note: A CPF adaptor fitting is not included with 15 000 psi /1000 bar /100MPa APM models. The module connector is compatible with 1/4 medium pressure tube system, HIP LM4 and LF4 Series, Autoclave Engr SF250CX Male and Female Series. See our CPF Data Sheet for additional adapter fittings.

Operation

ADVANCED PRESSURE MODULE (APM) INSTRUCTIONS

Pressure Connection

Crystal CPF System: Medium Pressure Female (MPF) (1/4" medium pressure tube system with 7/16-20 threads). See our CPF Brochure for further information.

CPF o-ring size and material: AS568A-012, Viton 80 durometer (P/N 3981).

For most applications CPF Fittings can be hand tightened for use up to 10 000 psi / 700 bar / 70 MPa (no tools required). Wrench tightening is recommended (to achieve a metal to metal cone seal) for applications where chemical compatibility of the process fluid and the o-ring are a concern, or for pressures above 10 000 psi / 700 bar / 70 MPa. We recommend a tightening torque of 120 in-lbs ±20 in-lbs. Please note this is only a fraction of the typical torque required to seal a 1/4" NPT fitting. If a torque wrench isn't practical to use, the fittings can be assembled as follows: Hand tighten fitting fully until the cone has bottomed out. Tighten an additional 20° using a wrench. Apply a small amount of media-compatible lubricant to the gland threads and male cone to increase fitting life, reduce the likelihood of galling, and promote sealing.

CAUTION: To achieve CPF maximum allowable working pressures no o-ring substitutions are allowed. See our CPF brochure and CES-003 CPF Safety Guide available from the website at crystalengineering.net for further detail.

WARNING: Pressurized hoses and associated equipment are potentially dangerous. Slowly bleed off pressure from the system being pressurized prior to connecting or disconnecting the Crystal APM.

Measuring Vacuum

All ranges of the APM can be used to measure moderate vacuum. Ranges 300 psi/30 bar/3 MPa and below are calibrated for vacuum use.

When measuring pressure less than ambient barometric conditions, a minus sign (-) will appear.

CAUTION: The APM is not recommended for continuous use at high vacuum.

Overpressure Conditions

The APM will read pressure up to approximately 110% of the rated pressure range. All JOFRA indicators will display overpressure warnings. For example, above 110% of the range, "OL" will display in red on an ASC-400, indicating an Overload Alarm. The zero function does not affect when the "OL" is displayed to indicate overpressure. So depending on the zero value, it is possible that the display will indicate "OL" without the maximum pressure being displayed.

For instance, if a 100 psi APM is zeroed when 30 psi is being applied, it will indicate that the overpressure condition has been reached at 80 psi. (i.e., 110% x 100 psi – 30 psi = 80 psi).

Overpressure can affect accuracy, but the effect is only temporary unless the sensor has been damaged. See <u>Specifications</u> for maximum allowable overpressure ratings.

Specifications

Modules

	Module Range	0 - 30% Gauge Full Scale	30 - 110% Gauge Full Scale		
	psi	± (% Full Scale)	± (% of Reading)		
	30	0.0075%	0.025%		
	100	0.0075%	0.025%		
psi	300	0.0075%	0.025%		
	1000	0.015%	0.05%		
	3000	0.015%	0.05%		
	10000	0.03%	0.1%		
	15000	0.03%	0.1%		

	bar	± (% Full Scale)	± (% of Reading)		
	3	0.0075%	0.025%		
	10	0.0075%	0.025%		
bar	30	0.0075%	0.025%		
	100	0.015%	0.05%		
	300	0.015%	0.05%		
	700	0.03%	0.1%		
	1000	0.03%	0.1%		

	kPa / MPa	± (% Full Scale)	± (% of Reading)		
	300	0.0075%	0.025%		
	1	0.0075%	0.025%		
kDa /	3	0.0075%	0.025%		
kPa/ MPa	10	0.015%	0.05%		
	30	0.015%	0.05%		
	70	0.03%	0.1%		
	100	0.03%	0.1%		

Accuracy (Gauge)

0 to 30% of Full Scale±(0.0075% of Full Scale), ±(0.015% of Full Scale), or ±(0.03% of Full Scale)

30 to 110% of Full Scale±(0.025% of Reading), ±(0.05% of Reading), or ±(0.1% of Reading)

VacuumFor 100 psi/10 bar/1 MPa and lower:
±(0.06% of Full Scale*)

For 300 psi/30 bar/3 MPa:
±(0.06% of Full Scale*) ±1 LSD

*Full Scale = -14.5 psi / -1.0 bar / -100.0 kPa.

Note: Exposure to environmental extremes of temperature, shock, and/or vibration may warrant a more frequent recertification period.

Accuracy specifications include all effects of linearity, hysteresis, repeatability, temperature, and stability for one year.

APM modules must be exercised and re-zeroed whenever exposed to significant changes in environmental conditions to achieve these specifications. To exercise a module, cycle the module between zero (ambient barometric pressure) and the pressure of interest. A properly exercised module will return to a zero reading (or return to the same ambient barometric reading).


CAUTION: Advanced Pressure Modules (APM) are not recommended for continuous use at high vacuum.

Accuracies, Ranges, and Resolutions

	psi	bar	kPa/MPa	Overpressure	psi	kg/cm2	inHg	inH20	mmHg	mmH20	kPa	bar	mbar	MPa
	30PSI			3.0 x	0.001	0.0001	0.001	0.01	0.01	1	0.01	0.0001	0.1	
0.025%		3BAR		3.0 x	0.001	0.0001	0.001	0.01	0.01	1	0.01	0.0001	0.1	
of Reading			300KPA	3.0 x							0.01	0.0001	0.1	
modules	100PSI			2.0 x	0.001	0.0001	0.01	0.1	0.1	1	0.01	0.0001	0.1	0.00001
		10BAR		2.0 x	0.001	0.0001	0.01	0.1	0.1	1	0.01	0.0001	0.1	0.00001
			1MPA	2.0 x							0.01	0.0001	0.1	0.00001
	300PSI			2.0 x	0.01	0.001	0.01	0.1	0.1		0.1	0.001	1	0.0001
		30BAR		2.0 x	0.01	0.001	0.01	0.1	0.1		0.1	0.001	1	0.0001
			3MPA	2.0 x							0.1	0.001	1	0.0001
	1KPSI			2.0 x	0.1	0.001	0.1				0.1	0.001		0.0001
0.05%		100BAR		2.0 x	0.1	0.001	0.1				0.1	0.001		0.0001
of Reading			10MPA	2.0 x							0.1	0.001		0.0001
modules	3KPSI			1.5 x	0.1	0.01	0.1				1	0.01		0.001
		300BAR		1.5 x	0.1	0.01	0.1				1	0.01		0.001
			30MPA	1.5 x							1	0.01		0.001
	10KPSI			1.5 x	1	0.01					1	0.01		0.001
0.1% of Reading		700BAR		1.5 x	1	0.01					1	0.01		0.001
			70MPA	1.5 x							1	0.01		0.001
modules	15KPSI*			1.3 x	1	0.01					1	0.01		0.001
		1000BAR*		1.3 x	1	0.01					1	0.01		0.001
			100MPA*	1.3 x							1	0.01		0.001

Resolutions shown are the maximum resolutions available. The resolution of your JOFRA device may be different.

▶ Ordering Information



SAMPLE PART NUMBERS

^{*} CPF adaptor fitting is not included. 1/4" medium pressure tube system is compatible with HIP LM4 and LF4 Series, Autoclave Engr SF250CX Male and Female Series. See our CPF data sheet for additional adapter fittings.

Pressure Conversions

Pressure conversion factors will vary, and be dependent on the JOFRA device connected to the APM. Refer to the documentation for your particular JOFRA device for its specifications.

Media Compatibility

Liquids and gases compatible with sensor and CPF fitting system. \\

Wrench-tight316 Stainless Steel

Finger-tight......316 Stainless Steel and Viton (internal CPF o-ring)

Communications

The APM interface uses a command – query – response protocol. Using this protocol the user can access the pressure measured by the module, as well as control and monitor the status of all the essential functions of the module.

INTERFACE DESCRIPTION

The APM can be remotely controlled using a PC terminal or computer program running the module in an automated system. A USB-based serial port connection is required for remote operation. The serial port connection also powers the module. Once connected, a user can write programs with Windows-based languages like Visual Basic.

SETTING UP THE SERIAL PORT FOR REMOTE CONTROL

A USB to serial cable (available with part number <u>APMCAL KIT</u>) is used for serial communications from the APM to a computer. To connect the module to a computer, attach it to a USB-to-serial cable, connected and installed on a PC. Connecting the module to the cable will automatically power it on.

Note: Be sure to set the appropriate COM port in your communications software: 9600 Baud; No parity; 8 Data bits; 1 Stop bit; No hardware flow control; Software flow control (XON/XOFF) recommended.

To verify the module has connected successfully, enter *IDN? within the communications program. This command will return information from the module.

USING COMMANDS

Command Types

The APM may be controlled using commands and queries. All commands may be entered using upper or lower case. Refer to Remote Commands and Error Codes for all available commands.

The commands are divided into the following types:

► Module Commands

Module Commands are used only by the module. For example, VAL? asks for the values displayed on the module display.

▶ Common Commands

Common Commands are standard commands used by most devices. These commands always begin with "*". For example, *IDN? tells the module to return its identification.

▶ Query Commands

Query Commands ask for Information, and always end with "?". For example, VAL? returns the current pressure reading of the module.

Command Processing

The data entered into the module is processed as follows:

- Data is taken as 7-bit ASCII.
- Characters are discarded if their decimal equivalent is less than 32 (space), except 10 (LF) and 13 (CR).
- The most significant data bit is ignored.
- Upper or lower case is acceptable.
- More than one command (Compound Commands) may be entered on one line. Commands are separated by semi-colons. For example,
 PRES UNIT KPA; FAULT?

Response Data Types

The data returned by the module can be divided into four types:

Integer

For most computers and controllers, integers are decimal numbers ranging from -32768 to 32768. For example, FAULT? could return an integer of 117. Refer to the Error Codes table for more information on error codes.

Floating

Floating numbers have up to 5 significant figures and exponents. For example, VAL? might return 25.345 PSI.

► Character Response Data (CRD)

Character Response Data is data returned as keywords. For example, PRES_UNIT? might return PSI.

► Indefinite ASCII (IAD)

Indefinite ASCII are any characters followed by a terminator. For example, *IDN might return CRYSTAL, APM003C, 678123, R130000.31/R08009.13.

Module Status

▶ Error Queue

If an error occurs due to invalid input or buffer overflow, its error code is sent to a 15 entry error queue. An error code is read from the queue with the FAULT? command. When the error queue is empty, FAULT? returns 0. The error queue is cleared when power is reset or when the clear command *CLS is entered.

► Input Buffer

The module stores all received data in a 128 character input buffer, which accommodates varying character input speeds.. If the input command/symbol stream is too fast, the module will send the XOFF command when the input buffer is full. When the module is able to continue to receive commands, it will generate the XON command. The module can receive up to eight more characters after the XOFF command has been sent. If more characters are received, it will stop receiving, clear the input buffers, interrupt command execution, and flag the error 120 (Input Buffer overflow).

In order to avoid interruptions in the command execution process:

- Turn on software (XON / XOFF) flow control on the PC side.
- Avoid long command lines (lines more than 96 characters).
- Send the *CLR command before starting the communication procedure.
- Check the error buffer sending the FAULT? command after the communication procedure is finished.

REMOTE COMMANDS AND ERROR CODES

The following tables list all the commands, and their descriptions, accepted by the APM.

Common Commands

Command	Description
*CLS	(Clear Status) Clears the error queue.
*IDN?	(Identification Query) Returns the manufacturer, model number, and firmware revision of the module.
*RST	Resets the module to the power up state.

Module Commands

Command	Description	
DAMP	Turns Damp on or off.	
DAMP?	Returns if Damp is on or off.	
FAULT?	Returns the most recent error code.	
PRES_UNIT	Sets the pressure unit for the indicated display.	
PRES_UNIT?	Returns the pressure unit from the indicated display.	
VAL?	Returns the measured values.	
ZERO_MEAS	Zeros the pressure module.	
ZERO_MEAS?	Returns the zero offset of the pressure module.	
CAL_STORE	This command must be sent anytime a parameter has changed.	

Parameter Units

Units	Meaning
PSI	Pressure in pounds per square inch.
KG/CM2 or KGCM2	Pressure in kilograms per square centimeter.
INHG	Pressure in inches of mercury at 0°C.
INH2O4C	Pressure in inches of water at 4°C.
INH2O20C	Pressure in inches of water at 20°C.
INH2O60F	Pressure in inches of water at 60°F.
INH2O68F	Pressure in inches of water at 68°F.
MMHG	Pressure in millimeters of mercury at 0°C.
MMH2O4C	Pressure in millimeters of water at 4°C.
MMH2O20C	Pressure in millimeters of water at 20°C.
CMH2O4C	Pressure in centimeters of water at 4°C.
CMH2O20C	Pressure in centimeters of water at 20°C.
MH2O4C	Pressure in meters of water at 4°C.
MH2O20C	Pressure in meters of water at 20°C.
MSW	Pressure in meters of sea water.
FTH2O4C	Pressure in feet of water at 4°C.
FTH2O20C	Pressure in feet of water at 20°C.
FTH2O60F	Pressure in feet of water at 60°F.
FTSW	Pressure in feet of sea water.
BAR	Pressure in bar.
MBAR	Pressure in millibar.
KPA or KPAL	Pressure in kiloPascal.
MPA or MPAL	Pressure in MegaPascal.
TORR	Pressure in Torr.

Error Codes

Error Number	Error Description
117	Unknown command.
118	Bad parameter value.
120	Input Buffer overflow.
122	Output Buffer overflow.
200	Invalid index.

ENTERING COMMANDS

Commands for the APM may be entered in upper or lower case. At least one space must be entered between the command and parameter. All other spaces are optional. Almost all commands for the APM are sequential, and any overlapped commands will be indicated as such. This section will briefly explain each of the commands and describe their general use. The descriptions will include any parameters that may be entered with the command, as well as the output of the command.

Common Commands

▶ *CLS

Clears the error queue. Also terminates all pending operations. When writing programs, use before each procedure to avoid buffer overflow.

▶ *IDN?

Returns the manufacturer, model number, and firmware revision of the APM. For example, *IDN? will return AMETEK, APM 0, 1.00.

Module Commands

▶ DAMP

Enables a damping filter on the readings returned by the VAL? command. The filter is a first-order smoothing filter with 90% damping percentage. For example, DAMP ON, will turn the damping function on.

► DAMP?

Returns the current state of the damping function. For example, DAMP? will return 0N if the damping function is on.

► FAULT?

Returns the error code number of an error that has occurred. The command may be entered when a previous command did not do what it was intended to do. Refer to the Error Codes table for more information on error code numbers.

▶ PRES_UNIT

Used to set the pressure unit for the indicated display. For example, PRES_UNIT PSI will set the pressure unit to psi. Send CAL_STORE to save the new parameters. Refer to the Parameter Units table for a complete list of available pressure units.

▶ PRES_UNIT?

Returns the current pressure unit being used.

▶ VAL?

Returns the value of the current measurement on the upper and lower display. For example, if the APM is measuring 5 psi, then VAL? will return 5.000000E+00, PSI.

▶ ZERO MEAS

Zeros the APM. Send CAL_STORE to save the new parameters.

► ZERO MEAS?

Returns the zero offset of the APM.



CALIBRATION SOFTWARE

APMCAL provides an easy way to calibrate your APM CPF Series Pressure Module. With a USB to serial cable, a computer running Windows, an APM module, and the APMCAL software program, you can:

- Calibrate the APM with an easy-to-use Userspan Wizard.
- View and change the calibration date.
- View module details including model number, serial number, and firmware version.
- Update the firmware version of the APM.

Order part number **APMCAL KIT** to receive the program on CD, plus a USB to serial cable.

Certifications

CE Statement



We declare that the APM is in accordance with the Electromagnetic Compatibility Directive, and the Pressure Equipment Directive per our declaration(s).

PED EC Declaration of Conformity



CF

DECLARATION OF CONFORMITY

According to ISO/IEC 17050-1:2004

Crystal Engineering Corporation An AMETEK Inc. company Manufacturer's Name:

Manufacturer's Address: San Luis Obispo, CA 93401

Declares under sole responsibility that the product as originally delivered

Product Name: nVision Series Reference Recorder XP2i Series and XP2i-DD Series Digital Pressure Gauge

30 Series and IS90 Series Pressure Calibrator nVision Series Lab Reference M1 Series Pressure Gauge

CPF Series Fittings and Hoses APM Advanced Pressure Module

Complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

Pressure Equipment Directive 97/23/EC

Have been designed and manufactured to the following

Product and Pressure Range Description

Pressure gauges, calibrators, and The above listed pressure gauges, calibrators and reference recorders are designed and manufactured in accordance with applicable portions of Annex I, Essential Safety Requirements, and sound engineering practices. These pressure gauges or calibrators (classified as pressure accessories per guideline 1/6) have a volume (V) of less than

0.1 liter (Article 3, 1.1.(a) first indent, Group 1 fluids).

less than or equal to 200 bar (2 900 psig) are not subject to the essential requirements

of the directive 97/23/EC (PED, Annex I) will be classed as Sound Engineering Practice (SEP), and shall not have the CE mark applied.

> 200 bar (2 900 psig)

Maximum Allowable Pressure (PS) For pressure gauges, calibrators and reference recorders for use on gases or liquids above 200 bar (2 900 psig) on Class 1 & 2 gases or liquids, Crystal Engineering maintains a technical file in accordance with Annex III, Module A (internal production

control) when CE mark is required.

Note: CPF Hoses (MPH series) not for use with Group 1 fluids and gases

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. The unit complies with all essential requirements of the Directives.

moreta Division Vice President & Crystal Business Manager David K Porter, P.E.

Crystal Engineering Corporation, an AMETEK Inc. company 708 Fiero Lane, Suite 9, San Luis Obispo, CA 93401, USA +1 805 595 5477

28 April 2014

AMETEK Denmark A/S Gydevang 32-34, 3450 Allerød, Denmark

EMC EC Declaration of Conformity





DECLARATION OF CONFORMITY

According to ISO/IEC 17050-1:2010

Manufacturer's Name: Crystal Engineering Corporation An AMETEK Inc. company

Manufacturer's Address 708 Fiero Lane, Suite 9

San Luis Obispo, CA 93401

Declares under sole responsibility that the product as originally delivered

Complies with the essential requirements of the following applicable European Directives, and carries the CE marking

accordingly:

FMC Directive 2004/108/FC

Standard **EHSR** EN 61326-1:2013

And conforms with the following product standards:

Description Class Status EN 61326-1:2013 Radiated Emissions Class B Pass Harmonized EN 61326-1:2013 (EN 61000-4-2) Electrostatic Discharge Harmonized Criteria A Pass EN 61326-1:2013 (EN 61000-4-3) Radiated Immunity Criteria A Pass Harmonized

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. The unit complies with all essential requirements of the Directives

Division Vice President & Crystal Business Manager David K Porter, P.E.

Crystal Engineering Corporation, an AMETEK Inc. company 708 Fiero Lane, Suite 9, San Luis Obispo, CA 93401, USA +1 805 595 5477

AMETEK Denmark A/S

Gydevang 32-34, 3450 Allerød, Denmark +45 4816 8000

21 March 2014

Support

CONTACT US

USA

Crystal Engineering • California* Tel +1 (800) 444 1850 Fax +1 (805) 595 5466 crystal@ametek.com

Mansfield & Green • Florida Tel +1 (800) 527 9999 cal.info@ametek.com **United Kingdom**

Tel +44 (0)1243 833 302 jofra@ametek.co.uk

France

Tel +33 (0)1 30 68 89 40 general.lloyd-instruments@ametek.fr Germany

Tel +49 (0)2159 9136 510 info.mct-de@ametek.de

Denmark

Tel +45 4816 8000 jofra@ametek.com India

Tel +91 22 2836 4750 jofra@ametek.com

Singapore

Tel +65 6484 2388 jofra@ametek.com

China

Shanghai

Tel +86 21 5868 5111

Beijing

Tel +86 10 8526 2111

Guangzhou

Tel +86 20 8363 4768

jofra.sales@ametek.com.cn

If calling, have ready the model number, serial number, date of purchase, and reason for return. You will receive instructions for returning the device to us.

WARRANTY

Crystal Engineering Corporation warrants the APM (Advanced Pressure Module) to be free from defects in material and workmanship under normal use and service for one (1) year from date of purchase to the original purchaser. It does not apply to batteries or when the product has been misused, altered or damaged by accident or abnormal conditions of operation.

Crystal Engineering will, at our option, repair or replace the defective device free of charge and the device will be returned, transportation prepaid. However, if we determine the failure was caused by misuse, alteration, accident or abnormal condition of operation, you will be billed for the repair.

CRYSTAL ENGINEERING CORPORATION MAKES NO WARRANTY OTHER THAN THE LIMITED WARRANTY STATED ABOVE. ALL WARRANTIES, INCLUDING IMPLIED WARRANTIES OF MERCHANTABILITY OR FITNESS FOR ANY PARTICULAR PURPOSE, ARE LIMITED TO A PERIOD OF ONE (1) YEAR FROM THE DATE OF PURCHASE. CRYSTAL ENGINEERING SHALL NOT BE LIABLE FOR ANY SPECIAL, INCIDENTAL OR CONSEQUENTIAL DAMAGES, WHETHER IN CONTRACT, TORT OR OTHERWISE.

Note: (USA only) Some states do not allow limitations of implied warranties or the exclusion of incidental or consequential damages, so the above limitations or exclusions may not apply to you. This warranty gives you specific legal rights and you may have other rights which vary from state to state.

^{*}ISO 17025 accredited calibration lab. (A2LA #2601.01).



© 2014 Crystal Engineering Corporation

708 Fiero Lane, Suite 9, San Luis Obispo, California 93401-8701

